



**University of
Zurich^{UZH}**

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2014

Activity classification based on inertial and barometric pressure sensors at different anatomical locations

Moncada-Torres, A ; Leuenberger, K ; Gonzenbach, R ; Luft, A ; Gassert, R

Abstract: Miniature, wearable sensor modules are a promising technology to monitor activities of daily living (ADL) over extended periods of time. To assure both user compliance and meaningful results, the selection and placement site of sensors requires careful consideration. We investigated these aspects for the classification of 16 ADL in 6 healthy subjects under laboratory conditions using ReSense, our custom-made inertial measurement unit enhanced with a barometric pressure sensor used to capture activity-related altitude changes. Subjects wore a module on each wrist and ankle, and one on the trunk. Activities comprised whole body movements as well as gross and dextrous upper-limb activities. Wrist-module data outperformed the other locations for the three activity groups. Specifically, overall classification accuracy rates of almost 93% and more than 95% were achieved for the repeated holdout and user-specific validation methods, respectively, for all 16 activities. Including the altitude profile resulted in a considerable improvement of up to 20% in the classification accuracy for stair ascent and descent. The gyroscopes provided no useful information for activity classification under this scheme. The proposed sensor setting could allow for robust long-term activity monitoring with high compliance in different patient populations.

DOI: <https://doi.org/10.1088/0967-3334/35/7/1245>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-100456>

Journal Article

Accepted Version

Originally published at:

Moncada-Torres, A; Leuenberger, K; Gonzenbach, R; Luft, A; Gassert, R (2014). Activity classification based on inertial and barometric pressure sensors at different anatomical locations. *Physiological Measurement*, 35(7):1245-1263.

DOI: <https://doi.org/10.1088/0967-3334/35/7/1245>

Activity Classification based on Inertial and Barometric Pressure Sensors at Different Anatomical Locations

A Moncada-Torres¹, K Leuenberger¹, R Gonzenbach², A Luft²
and R Gasser^{1†}

¹ Rehabilitation Engineering Laboratory, Department of Health Sciences and Technology, ETH Zurich, Zurich, Switzerland

² University Hospital Zurich, Zurich, Switzerland

E-mail: arturomo@ethz.ch, lkaspar@ethz.ch, roman.gonzenbach@usz.ch, andreas.luft@usz.ch, gassertr@ethz.ch

Abstract. Miniature, wearable sensor modules are a promising technology to monitor activities of daily living over extended periods of time. To assure both user compliance and meaningful results, the selection and placement site of sensors requires careful consideration. We investigated these aspects for the classification of 16 activities of daily living (ADL) in 6 healthy subjects under laboratory conditions using *ReSense*, our custom-made inertial measurement unit enhanced with a barometric pressure sensor used to capture activity-related altitude changes. Subjects wore a module on each wrist and ankle, and one on the trunk. Activities comprised whole body movements as well as gross and dextrous upper-limb activities. Wrist-module data outperformed the other locations for the three activity groups. Specifically, overall classification accuracy rates of almost 93% and more than 95% were achieved for the repeated holdout and user-specific validation methods, respectively, for all 16 activities. Including the altitude profile resulted in a considerable improvement of up to 20% in the classification accuracy for stair ascent and descent. The gyroscopes provided no useful information for activity classification under this scheme. The proposed sensor setting could allow robust long-term activity monitoring with high compliance in different patient populations.

Keywords: accelerometers, activity classification, barometric pressure sensor, dextrous activities, gross activities, gyroscopes, inertial measurement unit, machine learning, stair climbing, upper-limb, wearable sensors, whole body activities.

Submitted to: *Physiol. Meas.*

† Correspondence address: ETH Zurich, Rehabilitation Engineering Laboratory, LEO B 9.1, Leonhardstrasse 27, 8092 Zurich, Switzerland

1. Introduction

Miniature wearable sensor modules are a promising technology to monitor activities of daily living (ADL) over extended periods of time (Nyan et al. 2004), especially for aspects such as activity levels (Bouten et al. 1997), changes in habits (e.g. mental/physical decline in the elderly population (Hao & Foster 2008)), or functional recovery (e.g. following orthopedic or neurological injury) outside the clinical environment (Uswatte et al. 2000). A variety of sensors attached at different anatomical locations have been used in activity classification studies under different settings (Bao & Intille 2004, Stikic et al. 2008, Bonomi et al. 2009, Atallah et al. 2010, Roggen et al. 2011). The investigated activities range from sedentary (e.g. eating or drinking, reading, working on a computer) to ambulatory activities (e.g. walking, running, cycling, mopping). Table 1 presents a non-exhaustive list of previous works along with the selection and placement of sensors and the achieved overall activity classification accuracy.

Despite acceptable accuracy rates, it can be seen that cases where higher accuracies are reported either covered a reduced number of activities, therefore restricting the applicability to the identification of ADL, or used a large number of sensor modules, reducing comfort and non-obtrusiveness for the user. Further, in many of these studies, accurate classification of stair walking, and even more so differentiating stair ascent from stair descent, has proven to be a challenge. Yet, in order to determine the overall activity level and independence of subjects/patients in their home environment, being able to identify and discriminate stair ascent and descent is a key aspect. First of all, stair ascent and descent are different tasks with distinct biomechanical properties, including energy expenditure, and even performance time (Riener et al. 2002, Verlaan et al. 2009). Stair climbing can be a useful indicator of autonomy during the early rehabilitation stages of discharged patients (Paolucci et al. 2008), and is routinely assessed in the clinic (e.g. 12-Step Stair Test). Furthermore, stair-climbing-based assessments have shown moderate-to-high correlation with test scores such as the Berg Balance Scale (BBS) and the Fugl-Meyer Lower Extremity Assessment (FMA-LE) (Ng et al. 2013). Finally, a detailed analysis of level walking and stair ascent/descent is needed to evaluate subject's function/dysfunction during daily activities (Coley et al. 2005). For example, the presence of stair ascent but absence of stair descent can be an indicator of fear of falling in the elderly population (Butler et al. 2009).

Some studies have tried to tackle this problem by grouping both activities in a single "stair climbing" activity (Bao & Intille 2004) or even by grouping these activities together with level walking (Foerster et al. 1999). The works presented in Table 1 considered stair ascent and descent as individual activities and explicitly reported a classification accuracy for these activities, which is, in most cases, well below the overall classification accuracy.

There are currently no guidelines as to the selection and optimal placement of sensors for such investigations. The optimal selection and placement site of sensors

as well as the contribution of different sensor signal features for the classification of ADL has received little interest and therefore presents room for improvement. Optimal placement of sensor modules could reduce the required number of modules, guaranteeing good performance and compliance while minimizing discomfort for the subject. Knowing specifically which sensors provide useful information for the classification of ADL could further allow for more compact and lightweight designs, as well as optimized power management schemes for an extended runtime.

This paper investigates the optimal placement site and sensor selection using *ReSense* modules, our previously developed sensor module (Leuenberger & Gassert 2011), for the classification of 16 ADL in 6 healthy subjects under laboratory conditions. *ReSense* incorporates a 6 degree-of-freedom (DOF) inertial measurement unit (IMU) and a barometric pressure sensors (BPS), which serves as an altimeter. Although BPSs have previously been used in combination with inertial sensors to successfully discriminate real fall events from ADL (Bianchi et al. 2010) or to evaluate energy consumption of different physical activities (Ohtaki et al. 2005), their specific contribution to the classification of ADL has not been investigated before. We found that the addition of the altitude information significantly increases the classification accuracy of practically all of the 16 selected ADL, specifically allowing a clear distinction between stair ascent and descent. We further show that the gyroscope information does not contribute to the classification of the selected activities. As gyroscopes are energetically the most demanding, advanced energy management could allow wearable sensor modules to acquire data for longer periods of time. We also show that these modules can optimally be worn on the wrists, maximizing user compliance.

The manuscript is organized as follows. Section 2 gives a detailed description of how the study was conducted. Section 3 presents the obtained results, which are analyzed and discussed in Sec. 4. Section 5 closes the paper with our overall conclusions together with insights on possible future work.

2. Materials & Methods

An overview of the materials and methods used in the present study is provided in Figure 1.

2.1. Subjects

6 healthy subjects (4 males, 2 females; mean age 50.33 ± 21.59 years old) volunteered for this study. The study was approved by the local ethics committee. All participants provided written informed consent before being included.

2.2. Setting

The measurements took place indoors in a room and in a hall under common, everyday environmental conditions (light, noise, etc.) at the Zurich Center for Ambulatory

Table 1. Classification accuracy of previous studies which explicitly report values for stair ascent and descent.

Ref.	Sensors on module (no.)	Placement (no. of modules)	No. of activities	Training (Testing)	Classifier	Overall acc. [%]	Ascent acc. [%]	Descent acc. [%]
¹	2D accelerometer (1), digital compass (1), light detector (1)	Waist (1)	3	Combined (Combined)	kNN	NA	83.30	95.80
²	3D accelerometer (1)	Hip (2)	4	Combined (Combined)	ICA	NA	83.00	84.00
³	2D accelerometer (2)	Shoulder (2), hip (2), elbow (2), ankle (2), wrist (2), knee (2)	5	User (User)	Naive Bayes	~85.00	~80.00	~80.00
		Wrist (1)	8			NA	~ 5.00	~45.00
⁴	3D accelerometer (1), light sensor (1)	Wrist (1)	6	Combined (Combined)	Decision trees (C4.5)	87.00	~55.00	~80.00
⁵	3D accelerometer (1)	Thigh (1)	6	Combined (Combined)	Neural networks	91.70	61.50	44.30
⁶	3D accelerometer (1)	Wrist (1), ankle (1)	7	Combined (User)	kNN	96.98	86.54	90.38
⁷	1D accelerometer (2)	Thigh (1)	7	NA	Konohen SOM & Markov models	NA	42.00	64.00
⁸	3D accelerometer (1)	Pelvis (1)	8	User (User)	Plurality voting	65.15	42.86	0.00
⁹	3D accelerometer (1)	Wrist (2), thigh (1), trunk (1)	17	Combined (Combined)	kNN	90.61	79.20	73.33

¹ Lee & Mase (2001); ² Mantyjarvi et al. (2001); ³ Kern et al. (2003); ⁴ Maurer et al. (2006);

⁵ Kwapisz et al. (2011); ⁶ Wang et al. (2012); ⁷ Van Laerhoven & Cakmakci (2000); ⁸ Ravi et al. (2005); ⁹ Pirttikangas et al. (2006).

Rehabilitation (ZAR). The room was previously prepared with all the necessary furniture (e.g. bed, chair, table) and items (e.g. toothbrush, knife, peeler) required for the procedure. The hall provided large obstacle-free spaces where the subjects could perform some activities more easily and naturally than in a reduced, confined space (e.g. walking, running, stair climbing). Both areas were close to each other, thus reducing unnecessary transfers as much as possible.

2.3. Data Collection

Data was acquired with *ReSense* modules (Figure 2), watch-sized, low-power IMUs previously developed at our laboratory, which incorporate a BPS. Data from the latter is used to calculate an absolute altitude value with a resolution of 0.5 m. The module measures $\varnothing 25 \times 10$ mm and weighs 10 g, including a 200 mAh battery and a microSD card. With its intelligent power management scheme, this module achieves a runtime of more than 24 h for continuous sampling of all sensors at a frequency f_s of 50 Hz. If the gyroscopes are deactivated, the runtime can be extended to over 12 days (290 h). These runtimes are further doubled with and intelligent online detection of active/inactive periods. A detailed description of the sensor module and its technical specifications can be found in (Leuenberger & Gassert 2011).

Each subject wore five *ReSense* modules to record movement on both the left

(RS_{LW}) and right wrist (RS_{RW}), chest (RS_C), and the left (RS_{LA}) and right ankle (RS_{RA}), as shown in Figure 3. Each module was enclosed in a rapid prototyped housing, which provided protection, and was fixed to the body with an elastic band with VELCRO® straps, thereby reducing artifacts that may be caused by loose placement (Mathie et al. 2004). The total weight of each module, including the housing and elastic strap, was 21 g. The modules were configured to have a theoretical f_s of 50 Hz. This was sufficient for the purpose of this study since it has been shown that most human movements occur in a frequency band between 0.3 and 3.5 Hz (Sun & Hill 1993), that 99 % of their energy is contained below 15 Hz (Antonsson & Mann 1985), and that frequencies within the range of human body movement below 20 Hz are sufficient to assess daily physical activity (Bouten et al. 1997).

Subjects were instructed to perform a series of 16 different activities, listed in Table 4, as naturally as possible. These activities were selected based on three different criteria. First, we wanted to include a wide variety of activities covering the three types of movements indicative of different activity types, respectively recovery phases, which we desire to distinguish, i.e. whole body movements (e.g. walking, running, stair ascent, stair descent), gross (e.g. brushing teeth, drinking) and dexterous (e.g. writing, cutting food, peeling a carrot) upper-limb activities. We also included activities that have potential clinical relevance. For example, the Box & Block Test (BBT) was included since it is a well-accepted and widely used assessment of unilateral gross manual dexterity (Desrosiers et al. 1994). Stair climbing is a useful indicator for the assessment of patients undergoing rehabilitation under different scenarios (Ng et al. 2013, Paolucci et al. 2008). Finally, we also considered activities that have been previously used to test the performance of ADL monitoring systems, such as walking, standing, and laying down (Bao & Intille 2004, Maurer et al. 2006, Pirttikangas et al. 2006).

At the beginning and at the end of each data collection session (immediately before fixing the modules on the subjects and immediately after removing them), the five modules were shaken eight times together and in the same direction in order to generate a *shaking-marker event* (SME). SMEs would produce characteristic, notable peaks in the accelerometer recordings which would allow precise synchronization of the sensor data from the *ReSense* modules for post-experiment analysis.

All the performed activities (including the SMEs) were filmed with a video camera in order to obtain an objective, unbiased time-stamped reference of the movements performed by the subjects. In average, a data collection session lasted 11.24 ± 0.79 minutes, for a total of 67.46 minutes of data.

2.4. Preprocessing

Data from the *ReSense* modules were transferred from the integrated microSD card to a PC for post-experiment offline analysis. The raw sensor signals required several preprocessing steps, which are described in the following.




Figure 1. Overview of the materials and methods used in this study. Data were collected from six subjects each wearing five *ReSense* modules and were transferred to a computer for offline analysis via the integrated microSD card. The sensor signals were preprocessed before computing a set of predefined features. A reduced group of features was chosen based on the weights assigned by the ReliefF feature selection algorithm. Data were partitioned in different ways, according to the chosen validation method, and fed to the kNN classifier in the form of testing and training data sets. Finally, the performance of the system was evaluated using a confusion matrix. From the feature calculation stage, the solid arrow represents the original data set comprising all modules and all sensors (validated under the repeated holdout protocol), the dashed arrow represents the wrist module data for all sensors, and the dotted arrow represents the wrist module data excluding information from the BPS (these last two were validated additionally under the user-specific and leave-one-subject-out protocols).




Figure 2. *ReSense* module with labels for the main components. The module measures $\varnothing 25 \times 10$ mm and weighs 10 g, including a 200 mAh battery and a microSD card.

2.4.1. Resampling Due to variability in the accuracy of the onboard clock, sampling was not performed at exactly 50 Hz. In order for the signals to be further analyzed (segmented), they needed to contain the same amount of information (i.e. number of samples) per second. Sensor signals were resampled at 50 Hz with the help of the SMEs and the video recordings, matching the acceleration peaks. The real f_s was calculated by dividing the number of samples by the elapsed time. Resampling was done with an anti-aliasing finite impulse response (FIR) lowpass filter (LPF) with a Kaiser window of $\beta = 25$ (chosen empirically).

2.4.2. Filtering Each of the sensor signals recorded by the *ReSense* modules required a particular filtering stage. The triaxial accelerometer signals were first passed through a median filter with size $n = 3$ in order to remove any noise spikes (Mathie et al. 2002). The resulting signals contain the acceleration components due to gravity (a_g) and the acceleration components due to body movement (a_{bm}) linearly combined. In order to separate them, the signals were passed through an infinite impulse response (IIR) fourth order elliptic LPF with a cut-off frequency f_c of 0.3 Hz, a passband ripple of 0.01 dB, and a minimum stopband attenuation of 100 dB. Forward and time-reversed filtering was performed to avoid phase distortion (Mitra & Kuo 1998), thereby resulting in an overall eighth order LPF with a passband ripple of 0.02 dB and a minimum stopband attenuation of 200 dB. The lowpass filtered signal $acc_{\{x,y,z\}posture}$, which contains the a_g related to posture and orientation, was subtracted from the median filtered signal to yield $acc_{\{x,y,z\}active}$, which contains the a_{bm} related to activity and motion (Karantonis et al. 2006). This process is illustrated in Figure 4.

Each of the triaxial gyroscope signals was passed through a highpass filter (HPF)




Figure 3. Placement site and orientation of the five *ReSense* modules worn during the experiment.




Figure 4. Filtering and decomposition of the raw acceleration signals into their postural and active components, the first containing information related to posture and orientation and the latter containing information related to motion and activity. Forward and time-reversed low-pass filtering guarantee a phase-distortion free signal.

with the same specifications as the previous LPF used for the accelerometer data, including the f_c . This was done in order to remove any possible undesired low frequency components inherent to gyroscope measurements (e.g. drift). Forward and time-reversed filtering was also performed.

The altitude signal was passed through an IIR second order Butterworth LPF with a f_c of 0.07 Hz. Again, forward and time-reversed filtering was performed, resulting in an overall fourth order filter. These specifications were chosen empirically. As only the relative altitude change provides useful information regarding the altitude of the module, the DC component was removed. This was achieved by subtracting the arithmetic mean of the signal from itself.

Figure 5 shows an example of the filtered accelerometer, gyroscope, and BPS signals during a stair ascent/descent/ascent phase.

2.4.3. Fragmentation Each filtered signal was divided into smaller time segments using a sliding-window. With this method, the signal was fragmented into windows of fixed length with no inter-window gaps (Preece, Goulermas, Kenney, Howard, Meijer & Crompton 2009). This approach was selected due to the fact that the wide variety of activities performed by the subjects made it difficult to apply an event- or activity-based approach and because it is simple to implement. A window size N of 128 samples was chosen, representing 2.56 s of data with the given f_s . This window length has been shown to provide better recognition rates than other window sizes (Preece, Goulermas, Kenney & Howard 2009, Wang et al. 2007), allows capturing at least one complete cycle of periodic activities (e.g. a step while walking or stair climbing), and enables high-speed FFT calculations performed later. An overlap of 50% (i.e. 64 samples or 1.28 s) between windows was used, which has proven successful in previous works (Bao & Intille 2004, Preece, Goulermas, Kenney & Howard 2009).

2.4.4. Labeling Following fragmentation, each window was tagged with a label corresponding to the performed activity. This was done offline by matching the beginning and end of the window with the video recordings by visual inspection, using the SMEs as references in both.

It is important to note that all the data streams of all the modules were labeled according to the activity that was being performed by the subject at a given time and not by the activity that was performed by each body part. This greatly simplified the labeling task, which is usually a very time-consuming part in this type of studies. For example, if the subject performed the BBT, sensor data from the wrists, chest, and ankles were labeled as *BBT*, even though the activity of the chest and ankles resembled *standing*.

In the end, a total of 17 different labels were applied: one for each activity plus an additional one for the transitions in between activities (including the windows which




Figure 5. Filtered accelerometer, gyroscope, and BPS signals during a stair ascent/descent/ascent phase. The shown accelerometer and gyroscope signals correspond to the x axis.

captured the SMEs). Although several previous works have focused on activity/postural transitions (Wong & Wong 2008, Najafi et al. 2002, Godfrey et al. 2011), this is beyond the scope of this study. Therefore, for further analysis, windows labeled as *transition* were discarded, leaving only the labels corresponding to the 16 activities.

2.5. Feature Calculation

For each labeled window, a group of features was calculated that characterizes the data from the different sensors. Features in both time and frequency domain (the latter in the magnitude spectrum) were calculated. The features for the inertial sensors were selected based on those previously suggested in activity classification studies (Curone et al. 2010, Ravi et al. 2005, Maurer et al. 2006, Bouten et al. 1997, Karantonis et al. 2006, Herren et al. 1999, Baek et al. 2004, Bao & Intille 2004, Stikic et al. 2008). The features for the BPS (i.e. for the altitude profile), were chosen empirically with the objective of describing the potential altitude variations as well as possible. The complete feature set for each sensor is shown in Table 2. Note that different features were extracted from $acc_{\{x,y,z\}posture}$ and from $acc_{\{x,y,z\}active}$.

2.6. Data Subsets

Based on the original data set, which included information from all modules and features from all sensors, two additional subsets were extracted. The first subset included data from all the sensors of the wrist modules only. The second subset is just like the first, except that all the features from the altitude sensor were discarded. This was done in order to investigate the optimal sensor selection and placement site.

The original set, the complete wrist subset, and the altitude-discarded wrist subset are represented in Figure 1 by solid, dashed, and dotted arrows, respectively.

2.7. Feature Selection

The overall feature set comprised 131 different features. Besides having high dimensionality, which could result in computational performance issues, the original set may contain redundant or irrelevant features, which could negatively affect the results of subsequent analysis. Therefore, it was necessary to define a reduced feature set. This new set should have a high discriminative ability, i.e. it should show little variation between the same activities across different subjects but considerable variation between different activities (Preece, Goulermas, Kenney, Howard, Meijer & Crompton 2009).

Feature selection was done using the ReliefF algorithm (Kononenko 1994), which in our previous experience has proven to be effective. This method is an improved version of the original Relief algorithm proposed in (Kira & Rendell 1992), since it is more robust and can deal with incomplete and noisy data. The basic idea is to calculate the quality of attributes, selecting those that maximize distinction between classes. The algorithm randomly selects an instance, searches for K nearest neighbors from the

Table 2. Calculated features for each of the sensors integrated into *ReSense*.

Domain	Feature	Acc. (post.)	Acc. (act.)	Gyro.	Alt.
Time	Mean	✓			
	SD	✓	✓	✓	✓
	Variance	✓	✓	✓	✓
	IQR	✓	✓	✓	✓
	Perc. {3, 10, 20, 97}	✓	✓	✓	✓
	P2P amplitude	✓	✓	✓	✓
	Mean P2P amplitude	✓	✓	✓	
	Excessive kurtosis	✓	✓	✓	
	Slope				✓
	RMS	✓	✓	✓	✓
	SMA		✓		
	XY correlation		✓		
	YZ correlation		✓		
	XZ correlation		✓		
Frequency	Max. freq. component		✓		
	Energy		✓	✓	
	Entropy		✓	✓	
	Excessive kurtosis		✓		

SD: standard deviation, IQR: inter-quartile range, Perc. {3, 10, 20, 97}: percentiles 3, 10, 20, and 97, P2P: peak-to-peak, RMS: root mean square, SMA: signal magnitude area.

same class (called nearest hit H) and for K nearest neighbors from each of the other classes (called nearest miss M). It then updates a quality estimation $W[A]$ for all A attributes according to the values of the chosen attribute and the averaged contribution of H and M . The assigned weight ranges from -1 to 1, with large positive values given to important attributes (Robnik-Sikonja & Kononenko 2003). The algorithm iterates through all of the attributes. In this case, a value of 10 was chosen for K , as suggested in (Kononenko 1994).

Finally, a subset of the positively weighted attributes was selected. This was done by setting an empirical threshold, which was calculated as the average of the positive-valued attribute weights. All the features with a below-threshold weight were discarded.

2.8. Training & Testing Sets

Following the reduction of the feature set, it was necessary to separate the data into training and testing sets in order to perform a robust performance evaluation and make the most out of the data. Three protocols were chosen, which are explained in the following.

2.8.1. Repeated Holdout The data of all the subjects were put together into a big pool, randomly sampled and separated into a training data set and a testing data set with a proportion of two-thirds and one-third, respectively (Witten & Frank 2011). It is important to note that even though random sampling was performed, a proper class representation in both sets was guaranteed.

2.8.2. User-specific The data of each subject were put together as a data pool. Then again, random sampling with proper class representation was performed to generate a training data set and a testing data set with a proportion of two-thirds and one-third for each subject, respectively (Witten & Frank 2011).

2.8.3. Leave-one-subject-out The data of all subjects except for one were put together as the training set. The data of the remaining subject were put together as the testing set. Note that in this case there is no need to perform random sampling.

2.9. Classification

In all the validation protocols, the training and testing data sets served as inputs to the classification scheme. We chose the k -Nearest Neighbor (kNN) algorithm for classification, since it is straightforward, flexible, simple to implement, and its good performance has been demonstrated in previous works (Foerster et al. 1999, Preece, Goulermas, Kenney & Howard 2009, van den Berg-Emons et al. 2000). It is an instance-based learning scheme in which a multidimensional space is constructed. In this space, each dimension corresponds to one feature. The features are normalized (usually to a range between -1 and 1) in order to avoid distortion caused by disparity in the data range for each dimension (Duda et al. 2003). This space is then populated with training points. A testing point is then placed in this space and its class is predicted according to the majority of the k -nearest neighbors of training data (Preece, Goulermas, Kenney, Howard, Meijer & Crompton 2009). The algorithm can be tuned by choosing different values of k . Typically, k varies from 1 to a small percentage of the training data and must not be a multiple of the number of classes (Theodoridis & Koutroumbas 2003). Different metrics can be used to calculate the distance to the neighbors, e.g. Euclidean, the Manhattan or city-block, kD-tree, and ball tree metrics (Witten & Frank 2011). In our case, the best results were obtained for $k = 1$ with an Euclidean distance metric.

The classifier outputs a list of labels that predict each of the activity labels of the testing data. For the repeated holdout protocol, the classifier was run 10 times in order to reduce the effects of random sampling. For the user-specific protocol, 10 iterations were performed per subject (for the same reason), for a total of 60 iterations. In the case of the leave-one-subject-out protocol, only 6 iterations were run, i.e. one per subject. There was no need to perform several iterations in the latter, since no random sampling was involved.

Classified as →		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P
Walk	= A	100	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
Run	= B	0	100	0	0	0	0	0	0	0	0	0	0	0	0	0	0
Stair descent	= C	0	0	95	5.3	0	0	0	0	0	0	0	0	0	0	0	0
Stair ascent	= D	0	0	0	98	0	0	0	0	0	0	0	2.1	0	0	0	0
Stand	= E	0	0	0	0	90	1.6	0	1.6	0	0	0	1.6	0	4.8	0	0
Sit	= F	0	0	0	0	0	82	3	0	0	3	0	0	0	3	3	6.1
Lie down	= G	0	0	0	0	0	9.5	76	0	0	4.8	4.8	0	0	0	0	4.8
B&B Test	= H	0	0	0	0	0	0	0	96	0.9	1.8	0.9	0	0	0	0	0
Brush teeth	= I	0	0	0	0	0	0	0	0	89	3.6	3.6	0	0	0	0	3.6
Don/doff jacket	= J	0	0	0	0	0	0	4.9	0	0	85	2.4	2.4	0	4.9	0	0
Drink	= K	0	0	0	0	0	0	0	0	7.1	0	86	7.1	0	0	0	0
Butter bread	= L	0	0	0	9.1	0	0	0	0	0	9.1	0	82	0	0	0	0
Cut food	= M	0	0	0	0	0	0	0	0	4.5	0	0	0	95	0	0	0
Don/doff shoe	= N	0	0	0	0	0	0	0	18	0	0	0	0	0	82	0	0
Peel carrot	= O	0	0	0	0	0	5	0	0	0	0	0	0	0	0	95	0
Write	= P	0	0	0	0	0	6.7	0	0	0	0	0	0	0	0	3.3	90

Figure 6. Sample confusion matrix obtained for one iteration of the repeated holdout protocol in the complete wrist data subset. Two-thirds of the total data were included in the training set, while the remaining third was included in the testing set. The diagonal (highlighted cells) indicates successful classifications, while the off-diagonals represent false classifications. Normalized results are shown for better understanding. In this case, a total classification accuracy of 92.50% was achieved.

2.10. Evaluation

To quantify and evaluate the performance of each iteration, no matter the data subset or validation method, a confusion matrix was used. The total accuracy was calculated as the sum of successful classifications (elements of the main diagonal of the confusion matrix) divided by the total number of performed classifications (sum of all the elements of the confusion matrix). The reported overall accuracy and its SD correspond to the arithmetic mean and SD of the total accuracies obtained from the corresponding iterations.

The data from all five *ReSense* modules of all subjects were grouped as a complete set (RS_{LW} , RS_{RW} , RS_C , RS_{LA} , RS_{RA}), wrist set (RS_{LW} , RS_{RW}), chest set (RS_C), and ankle set (RS_{LA} , RS_{RA}). Each of these groups was evaluated under the repeated holdout validation protocol for different activity subsets: whole body (walk, run, stair descent, stair ascent, stand, sit, lie down), gross upper limb (BBT, brush teeth, don/doff jacket, drink), dextrous upper limb (butter bread, cut food, don/doff shoe, peel carrot), and all activities together. Then, the best resulting subset was chosen for a deeper analysis under the user-specific and leave-one-subject-out protocols.

3. Results

A sample (normalized) confusion matrix for one iteration of the repeated holdout protocol for the complete data set from the two wrist modules is shown in Figure 6. The diagonal shows correctly classified activities, while the off diagonals show false classifications. In this case, an overall classification accuracy of 92.5% was achieved.

The activity classification accuracies for the considered module subsets are

Table 3. Activity classification accuracies for different module locations under the repeated-holdout validation protocol.

Considered Module	Whole body	Upper limb (gross)	Upper limb (dextrous)	All activities	Average
All modules	96.27 \pm 0.38	88.87 \pm 0.73	87.61 \pm 0.49	83.59 \pm 0.53	89.08 \pm 5.29
Wrists	97.01 \pm 0.31	98.10 \pm 0.67	96.86 \pm 1.01	92.78 \pm 0.37	96.18 \pm 2.33
Chest	84.61 \pm 1.21	94.78 \pm 1.88	92.28 \pm 2.42	83.66 \pm 1.52	88.83 \pm 5.53
Ankles	95.78 \pm 1.16	91.44 \pm 2.11	82.75 \pm 3.46	72.70 \pm 1.99	85.66 \pm 10.20

summarized in Table 3. With an average classification accuracy of 96.18%, the wrist module subset outperformed the rest. Therefore, this subset was chosen for a deeper analysis under the user-specific and leave-one-subject-out protocols. These results are shown in Table 4.

To evaluate the contribution of the different features of each sensor, we looked at the attributes used for the classification of the wrist module subsets after the feature selection using the ReliefF algorithm and the empirically set threshold. On average, the number of used features was reduced from 131 to 44.33 (66.15%).

In all of the cases, none of the gyroscope features were retained by the ReliefF algorithm, not even in the data subset where the altitude information was discarded and only accelerometer and gyroscope data were available. Taking a closer look at the assigned weights, we were able to identify the features with better discriminative ability for this case. For the complete wrist data subset, the top-ranked features (i.e. the features that occupied the first 10 to 15 places) were the correlation coefficients between the $acc_{\{x,y,z\}active}$ signals, as well as different dispersion measurements (e.g. RMS, SD, percentiles $\{3, 10, 20, 97\}$, and P2P) of the $acc_{\{x,y,z\}posture}$ and altitude signals. For the altitude-discarded wrist subset, the top-ranked attributes were practically the same, except of course for the altitude features.

4. Discussion

Classification results for the different module subsets under the repeated holdout protocol clearly show that the wrist modules outperform other possible combinations: they provide higher and more consistent accuracies. This is especially evident when considering the complete activity set, showing an improvement of almost 10 percentage points over the second best module subset. It is important to note that grouping modules from different locations (e.g. *All modules* subset) can actually decrease the classification performance by diminishing the discriminative power of the features: the same features for one activity might have quite different values depending on the module site. For example, for the activity *Peel carrot*, the SD of the $acc_{\{x,y,z\}active}$ signals of the wrist modules is very different from the SD of the same signals of the ankle modules.

Given their better performance, wrist data subsets were chosen to be analyzed in more detail. Classification was performed with the user-specific and the leave-one-subject-out protocols in addition to the repeated holdout. In the case where all sensor

Table 4. Performed activities and their classification accuracy for the wrist modules under different validation protocols.

Type	Activity	Classification Accuracy [%]					
		Repeated holdout		User-specific		Leave-one-subject-out	
		All sensors	Alt. excluded	All sensors	Alt. excluded	All sensors	Alt. excluded
Whole body	Walk	100.00	95.74	100.00	95.69	93.21	65.12
	Run	97.22	92.45	94.50	87.04	12.90	12.90
	Stair descent	92.83	83.33	96.02	90.56	92.45	26.42
	Stair ascent	96.56	76.39	96.22	85.93	68.61	8.03
	Stand	89.14	80.98	91.76	90.25	70.17	70.17
	Sit	85.89	76.72	94.71	92.33	31.63	22.45
	Lie down	81.94	55.87	85.25	78.44	50.79	34.92
Upper-limb (gross)	BBT	97.35	93.73	99.03	96.80	57.06	64.26
	Brush teeth	91.61	80.00	90.43	93.28	32.93	30.49
	Don/doff jacket	82.41	65.95	86.75	75.96	62.81	57.85
	Drink	89.55	87.41	98.44	88.52	55.00	60.00
Upper-limb (dextrous)	Butter bread	80.17	76.32	94.74	89.19	0.00	2.86
	Cut food	96.77	82.33	96.40	91.74	14.06	1.56
	Don/doff shoe	84.00	66.48	83.62	68.36	32.69	30.77
	Peel carrot	93.14	85.44	98.10	94.37	1.64	37.70
	Write	88.41	81.25	92.36	91.78	10.00	15.56
Iterations		10	10	60	60	6	6
No. of features		46	40	48	45	45	42
Overall accuracy		92.78	83.65	95.06	90.72	57.12	45.89
SD		0.37	0.66	5.62	7.72	12.69	12.39

information was used, including the altitude data, satisfactory results were achieved with an overall accuracy of almost 93% and above 95% for the repeated holdout and the user-specific protocols, respectively. Looking at the accuracy for each activity, it can be seen that the activities that had the worst classification rate were *lie down*, *don/doff jacket/shoe*, *sit*, and *butter bread* with an accuracy between 80–85% for the repeated holdout protocol and between 83–86% for the user-specific protocol (except for *butter bread* and *sit*, which improved up to almost 95%). We believe that this is because these activities show greater intersubject variability and because they were performed by the subjects over a short period of time (e.g. in the case of *lie down*, the subjects only remained on their back for around 8 to 10 seconds) or for only a few repetitions (*don/doff jacket/shoe* was only performed twice per subject). This results in a reduced number of windows of the respective activity and thus small training and testing data sets. The classifier therefore has only little data from which to learn and even fewer data to test. Furthermore, a relatively small number of misclassifications can generate big drops in the accuracy rate.

Nevertheless, some of the activities achieved a classification accuracy of or very close to 100%: *walk*, *cut food*, *BBT*, and *run*. Specifically stair climbing, both for ascent and descent, achieved an accuracy of almost 93 and 97%, respectively, which is far superior to those previously reported in literature (Table 1). We attribute this improvement to the altitude data, since this is the main difference between *ReSense* and other activity monitors/classifiers or devices used in previous works. This is further

confirmed by the drop in accuracy rate of 9.5 to over 20 percentage points for these two particular activities under the repeated holdout and from 5 to over 10 percentage points under the user-specific protocols if the altitude data is discarded. Furthermore, these activities are not the only ones that suffer from an accuracy rate drop when the altitude data is not considered. Discarding altitude information results in a drop of classification performance for all activities in both protocols (except for *brush teeth* in the user-specific protocol). This is also reflected in the overall accuracy, with a performance improvement of 4 to 9 percentage points in the repeated holdout and user-specific protocols when all the sensor information is used.

Besides having a significant positive impact on the classification accuracy, the BPS has some other qualities that make it a suitable component for wearable sensor modules. Its power consumption is comparatively low, making up approximately 15 % of the system's total power consumption (1.25 mA) if only the accelerometer is used and to only 2.5 % of the system's total power consumption (7.8 mA) if the gyroscopes are used as well. Furthermore, it does not considerably increase the size or cost of the complete module. With a footprint of ($5 \times 5 \text{ mm}^2$) and a cost of approximately \$5 USD, the BPS is comparable to an accelerometer or a gyroscope.

For the leave-one-subject-out validation method, the results were greatly decreased as a result of the lack of user-specific data in the training set. For example, the overall accuracy dropped to a poor 46% in the worst case, where no altitude information is considered. However, it is still evident that the altitude data improves classification for 9 of the 16 ADL. For the specific case of stair ascent and descent, including the altitude information increases the accuracy by 61 and 66 %, respectively. Including the altitude data not only boosts the classification accuracy of these activities, but also enhances the overall accuracy rate by more than 11 percentage points.

Investigating the differential contribution of the different sensors to the discrimination of activities, it is important to note that, ideally, in order to achieve the best performance possible, the feature subset must be chosen according to the activities to be classified. However, making this decision *a priori* might not always be possible (e.g. in a real-life setting, in which the exact set of activities is unknown). We found that the most relevant features for classification of ADL using *ReSense* are mainly those that quantify the correlation between the $acc_{\{x,y,z\}active}$ signals and those that quantify the dispersion of the $acc_{\{x,y,z\}posture}$ signals and the altitude profile. Besides, it is important to note that none of the gyroscope features ranked high enough to be considered for the classification, i.e. all the gyroscope features were discarded.

Finally, we have to acknowledge some limitations of this study. First of all, the total number of subjects was small. The results and conclusions drawn from this data would have to be confirmed on a larger population. Also, the data collection was performed in a controlled laboratory-like environment. Even though the setting was conditioned to resemble an everyday environment and although the subjects were instructed to perform the tasks as naturally as possible, there is a possibility that their activity was influenced by the knowledge that they were participating in a study and by the fact that they were

being observed by the experimenter and recorded on video at the same time. However, using a video camera to validate and match the performance provides an objective time reference and is exempt of a possible bias from the experimenter or from the subject when keeping an activity log.

5. Conclusions and Future Work

This paper investigated the optimal sensor selection and placement site for activity classification. We obtained high classification accuracies for all 16 ADL using data from only the two wrist modules under the repeated holdout and user-specific validation protocols, and demonstrated the usefulness of including a BPS in the sensor module.

Besides resulting in better performance, placing the sensor modules on the wrists provides several advantages over other locations. Compared to the chest, where module placement would require some type of support around the trunk, or to the ankles, where the subject would need to bend down to attach the modules, the wrists are an easy to access position for the subject and require minimal effort. Moreover, portable monitoring devices on the wrists are unobtrusive and might be perceived as an additional accessory to the subject's outfit (e.g. watch, bracelet), making them more socially acceptable and, in the end, more comfortable for the subject to wear. All these facts would maximize subject compliance. This can thus provide more detailed and reliable information about the behaviour of users in the home environment for unsupervised long-term monitoring. Additionally, this setting allows to distinguish between three important groups of general activities: whole body movements, gross, and dextrous upper limb activities. Furthermore, placing the sensor modules on the wrists is advantageous in terms of investigating specific conditions in a number of diseases. For example, this configuration allows a robust monitoring of activity and at the same time provides a direct way to assess and quantify laterality and to compare the movement quality of the impaired and non-impaired upper limb in stroke patients. It could also prove advantageous to monitor arm tremor in Parkinson's patients. Nevertheless, modules on the chest and ankles could still be helpful for other types of applications. For example, the chest module could help to assess the subject's balance or the trunk's posture while performing different activities; the ankle sensors would certainly provide useful information for a detailed gait analysis.

The classification results are very satisfactory and encouraging, with good competitive overall classification accuracies and significantly better rates for stair ascent and descent in the repeated holdout and user-specific validation methods when altitude data was included. This demonstrates that the BPS integrated in the *ReSense* modules provides crucial information for the classification of practically all the ADL included in this study. It is worth emphasizing that this study embraced a rather large number of ADL (a total of 16), including whole body movements (e.g. walking, running, stair ascent and descent) and both gross (e.g. BBT, donning/doffing a jacket, drinking) and dextrous (e.g. cutting food, peeling a carrot, writing) upper-limb activities.

We found that the data from the accelerometers and from the integrated BPS were the most valuable for the classification of ADL under different settings. The fact that none of the three gyroscopes provided useful information allows a drastic extension of the battery runtime of our *ReSense* modules to more than 24 days by simply putting them into sleep mode and using an intelligent online detection of active/inactive periods. This could be very valuable for the long-term monitoring mode of our module for healthy subjects as well as for different patient groups in the future. However, this does not mean that the gyroscopes are not useful. For example, prior works have used them to identify stand-sit-stand transitions (Najafi et al. 2002). We would like to explore using sensor fusion techniques on data from both inertial sensors (i.e. accelerometers and gyroscopes) in combination with additional information from a 3D magnetometer to help determine the precise position and orientation of the wrist and reconstruct the movement trajectory of the upper limb in order to quantify and assess a patient's performance during motor tasks.

The results of the present study encourage further improvements and additional experiments. For example, we are interested in exploring the performance of the system using only data from one wrist module (particularly, from the dominant hand) to determine if two modules are necessary to classify these activities, or if the measurements could be performed with only one module without a significant drop in performance. For future activity classification studies, we could begin with a reduced, predefined set of features, which we now know perform better. This may reduce computational power and time needed for this type of analysis, making it more suitable for a possible online use. We are also interested in finding a way to deal with the *transition* events, since these are inherent to human activities. Solving this could allow *ReSense* to be used for everyday conditions – in contrast to the controlled ones presented here – without a significant drop in the classification accuracy. These improvements would lead to an enhanced robustness of the activity recognition. Finally, future work will focus on applying a methodology similar to the one presented here in stroke patients, in order to track recovery and monitor the transfer of clinical therapy to ADL in the home environment.

Acknowledgments

This work was supported by the National Center of Competence in Research in Neural Plasticity and Repair (NCCR Neuro) of the Swiss National Science Foundation, as well as by a private foundation. The authors wish to thank Kathrin Campen for her support at the Zurich Center for Ambulatory Rehabilitation (ZAR) and all the subjects who volunteered for the study. The authors are members of the Neuroscience Center Zurich (ZNZ) and the Rehabilitation Initiative and Technology Platform Zurich (RITZ).

References

- Antonsson E K & Mann R W 1985 The frequency content of gait *Journal of Biomechanics* **18**(1), 39–47.
- Atallah L, Lo B, King R & Yang G Z 2010 Sensor for activity detection using wearable accelerometers in ‘Body Sensor Networks (BSN), 2010 International Conference on’ IEEE pp. 24–29.
- Baek J, Lee G, Park W & Yun B 2004 Accelerometer signal processing for user activity detection in ‘Knowledge-Based Intelligent Information and Engineering Systems’ Springer pp. 610–617.
- Bao L & Intille S 2004 Activity recognition from user-annotated acceleration data *Pervasive Computing* pp. 1–17.
- Bianchi F, Redmond S J, Narayanan M R, Cerutti S & Lovell N H 2010 Barometric pressure and triaxial accelerometry-based falls event detection *Neural Systems and Rehabilitation Engineering, IEEE Transactions on* **18**(6), 619–627.
- Bonomi A, Goris A, Yin B & Westerterp K 2009 Detection of type, duration, and intensity of physical activity using an accelerometer *Medicine & Science in Sports & Exercise* **41**(9), 1770–1777.
- Bouten C, Koekkoek K, Verduin M, Kodde R & Janssen J 1997 A triaxial accelerometer and portable data processing unit for the assessment of daily physical activity *Biomedical Engineering, IEEE Transactions on* **44**(3), 136–147.
- Butler A A, Menant J C, Tiedemann A C & Lord S R 2009 Age and gender differences in seven tests of functional mobility *Journal of NeuroEngineering and Rehabilitation* **6**, 31.
- Coley B, Najafi B, Paraschiv-Ionescu A & Aminian K 2005 Stair climbing detection during daily physical activity using a miniature gyroscope *Gait & posture* **22**(4), 287–294.
- Curone D, Bertolotti G, Cristiani A, Secco E & Magenes G 2010 A real-time and self-calibrating algorithm based on triaxial accelerometer signals for the detection of human posture and activity *Information Technology in Biomedicine, IEEE Transactions on* **14**(4), 1098–1105.
- Desrosiers J, Bravo G, Hebert R, Dutil E & Mercier L 1994 Validation of the Box and Block Test as a measure of dexterity of elderly people: reliability, validity, and norms studies *Archives of physical medicine and rehabilitation* **75**(7), 751–755.
- Duda R, Hart P & Stork D Pattern classification 2nd edn Wiley New York pp. 187–191.
- Foerster F, Smeja M & Fahrenberg J 1999 Detection of posture and motion by accelerometry: a validation study in ambulatory monitoring *Computers in Human Behavior* **15**(5), 571–583.
- Godfrey A, Bourke A, Ólaighin G, Van De Ven P & Nelson J 2011 Activity classification using a single chest mounted tri-axial accelerometer *Medical engineering & physics* **33**(9), 1127–1135.
- Hao Y & Foster R 2008 Wireless body sensor networks for health-monitoring applications *Physiological measurement* **29**, R27.
- Herren R, Sparti A, Aminian K & Schutz Y 1999 The prediction of speed and incline in outdoor running in humans using accelerometry *Medicine & Science in Sports & Exercise* **31**(7), 1053.
- Karantonis D, Narayanan M, Mathie M, Lovell N & Celler B 2006 Implementation of a real-time human movement classifier using a triaxial accelerometer for ambulatory monitoring *Information Technology in Biomedicine, IEEE Transactions on* **10**(1), 156–167.
- Kern N, Schiele B & Schmidt A 2003 Multi-sensor activity context detection for wearable computing *Ambient Intelligence* pp. 220–232.
- Kira K & Rendell L 1992 A practical approach to feature selection in ‘Proceedings of the ninth international workshop on Machine learning’ pp. 249–256.
- Kononenko I 1994 Estimating attributes: analysis and extensions of RELIEF in ‘Machine Learning: ECML-94’ Springer pp. 171–182.
- Kwapisz J R, Weiss G M & Moore S A 2011 Activity recognition using cell phone accelerometers *SIGKDD Explor. Newsl.* **12**(2), 74–82.
- Lee S W & Mase K 2001 Recognition of walking behaviors for pedestrian navigation in ‘Control Applications, 2001. (CCA ’01). Proceedings of the 2001 IEEE International Conference on’ pp. 1152–1155.

- Leuenberger K & Gassert R 2011 Low-Power Sensor Module for Long-Term Activity Monitoring *in* ‘Conference Proceedings of the IEEE Engineering in Medicine and Biology Society (EMBC)’.
- Mantyjarvi J, Himberg J & Seppanen T 2001 Recognizing human motion with multiple acceleration sensors *in* ‘Systems, Man, and Cybernetics, 2001 IEEE International Conference on’ Vol. 2 pp. 747–752 vol.2.
- Mathie M, Coster A, Lovell N & Celler B 2004 Accelerometry: providing an integrated, practical method for long-term, ambulatory monitoring of human movement *Physiological measurement* **25**, R1.
- Mathie M, Lovell N, Coster A & Celler B 2002 Determining activity using a triaxial accelerometer *in* ‘Engineering in Medicine and Biology, 2002. 24th Annual Conference and the Annual Fall Meeting of the Biomedical Engineering Society EMBS/BMES Conference, 2002. Proceedings of the Second Joint’ Vol. 3 IEEE pp. 2481–2482.
- Maurer U, Rowe A, Smailagic A & Siewiorek D 2006 Location and activity recognition using eWatch: A wearable sensor platform *Ambient Intelligence in Everyday Life* pp. 86–102.
- Mitra S & Kuo Y Digital signal processing: a computer-based approach McGraw-Hill pp. 224–226, 531–532.
- Najafi B, Aminian K, Loew F, Blanc Y & Robert P 2002 Measurement of stand-sit and sit-stand transitions using a miniature gyroscope and its application in fall risk evaluation in the elderly *Biomedical Engineering, IEEE Transactions on* **49**(8), 843–851.
- Ng S S, Ng H H, Chan K M, Lai J C, To A K & Yeung C W 2013 Reliability of the 12-step ascend and descend test and its correlation with motor function in people with chronic stroke *Journal of Rehabilitation Medicine* **45**(2), 123–129.
- Nyan M, Tay F, Koh T, Sitoh Y & Tan K 2004 Location and sensitivity comparison of MEMS accelerometers in signal identification for ambulatory monitoring *in* ‘Electronic Components and Technology Conference, 2004. Proceedings. 54th’ Vol. 1 IEEE pp. 956–960.
- Ohtaki Y, Susumago M, Suzuki A, Sagawa K, Nagatomi R & Inooka H 2005 Automatic classification of ambulatory movements and evaluation of energy consumptions utilizing accelerometers and a barometer *Microsystem technologies* **11**(8-10), 1034–1040.
- Paolucci S, Bragoni M, Coiro P, De Angelis D, Fusco F R, Morelli D, Venturiero V & Pratesi L 2008 Quantification of the probability of reaching mobility independence at discharge from a rehabilitation hospital in nonwalking early ischemic stroke patients: a multivariate study *Cerebrovascular Diseases* **26**(1), 16–22.
- Pirttikangas S, Fujinami K & Nakajima T 2006 Feature selection and activity recognition from wearable sensors *Ubiquitous Computing Systems* pp. 516–527.
- Preece S, Goulermas J, Kenney L & Howard D 2009 A comparison of feature extraction methods for the classification of dynamic activities from accelerometer data *Biomedical Engineering, IEEE Transactions on* **56**(3), 871–879.
- Preece S, Goulermas J, Kenney L, Howard D, Meijer K & Crompton R 2009 Activity identification using body-mounted sensors - a review of classification techniques *Physiological Measurement* **30**, R1–R33.
- Ravi N, Dandekar N, Mysore P & Littman M 2005 Activity recognition from accelerometer data *in* ‘Proceedings of the 17th conference on Innovative applications of artificial intelligence’ Vol. 3 AAAI Press pp. 1541–1546.
- Riener R, Rabuffetti M & Frigo C 2002 Stair ascent and descent at different inclinations *Gait & Posture* **15**(1), 32–44.
- Robnik-Sikonja M & Kononenko I 2003 Theoretical and empirical analysis of ReliefF and RReliefF *Machine learning* **53**(1), 23–69.
- Roggen D, Magnenat S, Waibel M & Troster G 2011 Wearable computing *Robotics Automation Magazine, IEEE* **18**(2), 83–95.
- Stikic M, Huynh T, Van Laerhoven K & Schiele B 2008 ADL recognition based on the combination of RFID and accelerometer sensing *in* ‘Pervasive Computing Technologies for Healthcare, 2008.

- PervasiveHealth 2008. Second International Conference on' pp. 258–263.
- Sun M & Hill J 1993 A method for measuring mechanical work and work efficiency during human activities *Journal of Biomechanics* **26**(3), 229 – 241.
- Theodoridis S & Koutroumbas K Pattern recognition 2nd edn Elsevier pp. 44–46.
- Uswatte G, Miltner W, Foo B, Varma M, Moran S & Taub E 2000 Objective measurement of functional upper-extremity movement using accelerometer recordings transformed with a threshold filter *Stroke* **31**(3), 662–667.
- van den Berg-Emons H, Bussmann J, Balk H & Stam H 2000 Validity of ambulatory accelerometry to quantify physical activity in heart failure *Scandinavian Journal of Rehabilitation Medicine* **32**(4), 187–192.
- Van Laerhoven K & Cakmakci O 2000 What shall we teach our pants? *in* 'Wearable Computers, 2000. The Fourth International Symposium on' pp. 77–83.
- Verlaan L, Senden R, Storken G, Heyligers I & Grimm B 2009 Outcome Assessment in Orthopaedics using a Stair Climbing Test and Accelerometer based Motion Analysis: a Feasibility Study *Journal of Bone & Joint Surgery, British Volume* **91**(SUPP III).
- Wang J S, Chuang F C & Yang Y T C 2012 A wearable physical activity sensor system: its classification algorithm and performance comparison of different sensor placements *in* 'Advanced Intelligent Computing Theories and Applications. With Aspects of Artificial Intelligence' Springer pp. 447–454.
- Wang N, Ambikairajah E, Lovell N & Celler B G 2007 Accelerometry based classification of walking patterns using time-frequency analysis *in* 'Engineering in Medicine and Biology Society, 2007. EMBS 2007. 29th Annual International Conference of the IEEE' IEEE pp. 4899–4902.
- Witten I & Frank E Data mining: Practical machine learning tools and techniques 3rd edn Morgan Kaufmann chapter 4, 5.
- Wong W Y & Wong M S 2008 Detecting spinal posture change in sitting positions with tri-axial accelerometers *Gait & Posture* **27**(1), 168 – 171.